



DECLARATION OF CONFORMITY

Product type : Thermal mass flow meter
Brand name(s) : *VP Instruments, VP FlowScope M*
Model numbers : VPM-RXXX-PXXX-DXXX
Where X can be any number or letter

Power supply : 12 – 24 VDC
Max pressure : 10 bar gage
Manufacturer : Van Putten Instruments B.V.
Buitenwatersloot 335
2614 GS Delft
The Netherlands

Year of first making : 2016


Date of issue : April 01-2016

The listed devices are in compliance with the council directives of the European Community and are marked with the CE-marking. The safety and installation requirements of the product documentation must be observed.

EMC-directive : 2014/30/EC
Emission : EN 60950-1 (2013) Class A *
EN 61326-1 (2007) Class A *
EN 61000-3-2 (2014) Class A *
EN 61000-3-3 (2013) Class A *
Immunity : EN 61326-1 (2013) Class A *

22-April-2013

Van Putten Instruments BV


Pascal F.A. M. van Putten
Managing Director

*including alterations